

Whitepaper

Maintaining MALDI-TOF Performance: Why Active Ion Optics Cleaning Matters

Authors: Dr. Diana Kalinina, Director Product Management, Bruker Microbiology & Infection Diagnostics
Andreas Haase and Markus Kayser, MALDI Development, Bruker Microbiology & Infection Diagnostics

Summary

MALDI-TOF MS performance is critically dependent on ion source cleanliness. Contaminated ion optics can reduce the precise acceleration and guiding of ions to the detector. This affects the achievable peak resolution and sensitivity, which has an impact on the reliability of identification. Contamination is an inherent result of the MALDI ionization processes, leading to gradual performance loss over time. While various approaches attempt to mitigate the effects, only active removal of contaminants ensures sustained analytical performance. Compensation approaches might reduce the negative effects of contaminants but only postpone the inevitably needed cleaning, while compromising on performance in the meantime. With over 30 years of MALDI-TOF experience, Bruker knows how to tackle this challenge and combines IR laser-based ion optics cleaning, optimized matrix quality and IDealTune™ autotuning to create a comprehensive performance management strategy for the MALDI Biotyper®.

Ion source contamination – The fundamental problem

Ion source contamination in MALDI-TOF systems is a universal physical and chemical process involving the deposition of matrix components, salts, organic residues, and other materials on ion source surfaces. These deposits accumulate on extraction electrodes, ion lenses, and surrounding components, altering electric fields and ion trajectories, having negative effects on overall system performance.

This contamination is inevitable due to the nature of MALDI ionization, where laser desorption produces plumes containing both ionized and neutral particles. Over time, this leads to measurable declines in sensitivity, reproducibility, and instrument stability. High-mass detection is particularly affected.

What are the consequences?

Considering the contamination built up over time, users will experience a decline in system performance, reducing the reliability of species identification. Laboratories will typically observe an increase in “no ID” results and a higher need for repeat measurements. This reduced analytical efficiency has clear operational and financial consequences. More frequent repeat analyses drive up consumable usage, increase the total number of samples processed, and may even necessitate the use of alternative identification methods. In all cases, additional staff time is required to obtain reliable results - raising overall operational costs.

Moreover, the situation can rapidly escalate into a self-reinforcing cycle. Higher sample throughput due to repeat testing accelerates contamination build up, further degrading system performance.

This negative spiral not only compromises data quality but also, leads to greater wear and tear on system components that need to compensate for the negative effects of contamination build-up.

Consequently, laboratories may face more frequent service interventions and unplanned downtime. In routine microbiology workflows - especially when processing time-critical samples - consistent system availability and optimal performance are essential. Ensuring that the instrument operates at peak condition every day is therefore not only a technical requirement, but a critical factor in maintaining laboratory efficiency and diagnostic confidence. The following section outlines how Bruker's approach effectively addresses these inherent technological limitations.



Figure 1
Visualization of system performance over time.
This figure shows the anticipated performance behaviors of MALDI-TOF systems with no contamination cleaning strategy (grey), a compensation strategy (orange) and Bruker's holistic three-layered approach (blue).

Bruker's approach: active contamination management

Bruker addresses contamination through a holistic three-layer strategy consisting of reduction, cleaning and stabilization (Fig.1). This approach ensures not only immediate recovery of performance but also long-term operational stability.

IR laser-based ion optics cleaning removes contamination directly from critical ion source surfaces. This solution targets the root cause by eliminating deposited material. This prevents long-term accumulation, preserves ion transmission efficiency, and stabilizes performance over extended operational periods.

Bruker's Perpetual Ion Source™ with IR-laser self cleaning removes built up matrix and sample contamination from source electrodes by vaporizing the residue with a focused IR laser beam, all without breaking the instrument's vacuum.

Optimal performance is just a few clicks and a couple of minutes away – no need to wait for a service appointment!

Matrix quality plays a decisive role in contamination dynamics. High-purity, application-specific matrices minimize contaminant introduction, improve crystallization behavior, and reduce background noise. This contributes to improved signal-to-noise ratios, enhanced reproducibility, and lower contamination build-up over time.

With over 30 years of experience in MALDI mass spectrometry, Bruker tested and compared almost all matrices and substitutes available on the market. None of them were satisfactory, leading to the development of an own process for the refining and purification of the matrix substances to remove potential sources of contamination. The result is Bruker Microbiology & Infection Diagnostics purified MALDI matrix with outstanding quality criteria. Examples of the quality criteria are purity, crystallization behavior, background noise and signal-to-noise ratio and more.

IDealTune provides automated system optimization, adjusting acquisition parameters to stabilize system performance. This ensures consistent performance over time, reduces service intervention, and enables reliable routine operation.

This three-layered holistic approach aims to provide the following user benefits:

- Stable system performance
- High system uptime
- Low cost of ownership
- Reduced manual maintenance

What happens inside a MALDI-TOF?

Contamination: During normal MALDI-TOF operation, laser desorption creates a plume that deposits microscopic layers of matrix, salts, and analyte especially onto proximate ion source apertures.

Sophisticated cleaning setup: When a cleaning cycle is initiated, the contaminated surfaces are irradiated by an infrared laser beam. The used wavelength was chosen to avoid influence of matrix contamination on the cleaning arrangement itself. This is an advantage to other approaches using wavelengths in the ultraviolet region.

Desorption: The IR laser beam is strongly absorbed by the contaminated ion optics surfaces. The localized heating causes the contamination to undergo a transition into the gas phase.

Storage: The evaporated material from the ion source precipitates mainly in an intended recipient zone before reaching the pumps. This reduces environmental impact.

Limitations of alternative approaches

Technologies such as ion deflection in MALDI-TOF systems operate exclusively at the level of ion trajectories in the gas phase. While they may improve spectral appearance by deflecting unwanted ions before the detector, they do not prevent contaminants from reaching the ion source, nor can they remove contaminants from the ion source (Fig.2). Contamination continues to accumulate over time, eventually requiring maintenance intervention. The lack of a self-cleaning ion source cannot be compensated by means to improve just the output, they do not equate to reduced cleaning requirements.

Additionally, such compensation strategies may put increased strain on other parts of the system. They may hide or reduce the visibility of negative effects of contamination on the results level, but to reach these results, other components will be subject to greater stress, creating more wear and tear, resulting in a higher cost of ownership or increased system downtime due to more frequent service interventions.

Bruker's approach



Figure 2
Setup of Bruker's approach.
Schematic of MALDI-TOF components. The orange area shows the occurrence of contamination risk. The blue arrowheads show Bruker's three-layered approach covering reduction of contamination by matrix purity (A), cleaning with infrared laser (B) and stabilization by IDealTune (C), compared to retroactive strategies like ion deflection (D) which do not address the root cause.

Conclusion

Strategies predominantly focusing on managing contamination effects through instrument design optimization may delay performance degradation, but do not eliminate contamination. Here Bruker's philosophy fundamentally differs by actively reducing and removing contaminants, resulting in superior long-term stability and reduced maintenance demands. The ion source cleaning approach is supported by a high-quality matrix, reducing possible contaminants, and an effortless autotuning concept helping to stabilize system performance over time by adjusting parameters (Fig.3).

The integration of reduction, cleaning, and stabilization translates into tangible user benefits: higher reliability, reduced downtime, improved consistency of results, and lower total cost of ownership. Laboratories benefit from predictable performance and minimized service interruptions.

Sustainable MALDI-TOF performance requires active contamination management. By combining IR laser-based cleaning, optimized consumables, and automated tuning, Bruker delivers a uniquely robust and reliable solution for long-term operation.



Figure 3
Schematic of Bruker's approach to maintain system performance.
The ideal way to maintain MALDI-TOF system performance is to approach it from several angles. This includes ion optics cleaning (cleaning), supported by optimized matrix quality (reduction) and completed by an autotuning process like IDealTune (stabilization).

No. 1931367 © 08-2026 Bruker Daltonics

Online information
bruker.com/microbiology

Bruker Daltonics GmbH & Co. KG
Bremen · Germany
Phone +49 (0) 421-2205-0

info.md@bruker.com

Bruker Scientific LLC
Billerica, MA · USA
Phone +1 (978) 663-3660